

IZJAVA O SKLADNOSTI *DECLARATION OF CONFORMITY*

(v skladu z dodatkom VII Direktive za medicinske pripomočke 93/42/EGS)
(in accordance with Annex VII of Medical device directive 93/42/EEC)

Podjetje/Company:

INTERDENT[®] d.o.o.

Naslov/Address:

Opekarniška cesta 26, SI - 3000 CELJE

Na lastno odgovornost izjavljamo, da sledeči proizvodi, razreda I (pravilo 5)
We herewith declare on our responsibility that the following Class I Products (rule 5)

MREŽICA ZA UČVRSTITEV AKRILATA / REINFORCING NET

REF 0352

ustrezajo bistvenim zahtevam Direktive o medicinskih pripomočkih 93/42/EGS.
comply with essential requirements of the Medical Devices Directive 93/42 EEC.

HARMONIZIRANI STANDARDI / HARMONISED STANDARDS:

ISO 13485:2003+AC 2007: Medicinski pripomočki – Sistem vodenja kakovosti-
Zahteve za zakonodajne namene / *Medical devices – Quality management systems – Requirements for regulatory purposes*

EN ISO 14971:2007 Medicinski pripomočki-Uporaba obvladovanja tveganja pri
medicinskih pripomočkih / *Medical devices - Application of risk management to medical devices*

EN ISO 10993-1:2003: Biološko vrednotenje medicinskih pripomočkov – 1.del: Ocena
in preskusi / *Biological evaluation of medical devices – Part 1: Evaluation and testing*

EN ISO 15223-1:2012: Medicinski pripomočki – Simboli za označevanje medicinskih
pripomočkov, označevanje in podatki, ki jih mora podati dobavitelj- 1.del: Splošne
zahteve / *Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – part 1: general requirements*

Celje, 24.03.2014
Place, Date

Anja Šraj, univ.dpl.chem.
Responsible person for MD and technical files


Signature: